



به نام خدا

سمینار دو روزه آشنایی مقدماتی با درمانهای
پیشرفته با دستگاه توموتراپی

کنترل کیفی در درمانهای IMRT شتابدهنده (در عبور به Helical Tomotherapy)

PARDISNOOR NILOO

سجاد میرزایی

Radiotherapy Chain:

1. Localization:

- Countoring Target and OAR
- Multimodality : Image registration


2. Dose prescription

3. Treatment plan optimization

4. Treatment delivery

Treatment device:

- Mechanical accuracy of the system
- Localization system




Delivery of dose within $\pm 5\%$

- Sources of uncertainty
 - Treatment planning (estimated uncertainty of the order of $\pm 2\%$)
 - Machine performance on the day ($\pm 2\%$)
 - Patient set-up and movement ($\pm 3\%$)

 - Absolute dosimetry/calibration
 - Relative dosimetry (%depth dose, profiles, output factors)

Not much room for error in dosimetry ...



Quality Assurance and Quality Checks


Task Group 142 report: Quality assurance of medical accelerators¹

ESTRO

- The goal of a QA program for linear accelerators is to assure that the machine characteristics do not deviate significantly from their baseline values acquired at the time of acceptance and commissioning*

***Task Group 142 report**

ISPPC



Quality assurance of equipment	
Equipment tests to be performed on a regular basis	
Dosimetry <ul style="list-style-type: none"> • Output • Beam symmetry/flatness • Energy • Dose linearity 	Similar to 3D conformal QA
Mechanical <ul style="list-style-type: none"> • Jaws/MLC/couch absolute and relative position and speed • Gantry/collimator/couch angle • Gantry/collimator/couch rotation iso-centre 	Much higher accuracy
Imaging <ul style="list-style-type: none"> • Imaging and treatment iso-centre coincidence • Reproducibility • Contrast, uniformity, noise and image quality • HU constancy • Portal imager dose 	Similar to 3D conformal QA (at least in principle)

Quality assurance of equipment

IMRT (linac)

- **MLC:** Position and speed

VMAT (linac)

- **MLC:** Position and speed
- **Gantry:** Position and speed
- **Dose rate**

Tomotherapy

- **Synchrony:** MLC opening, table position and gantry position

Cyber knife

- **Isocentre:** Advanced Winston Lutz test

Simulation of delivered plan by use of DynaLog files

Commissioning of TPS

- **Measurement of leaf gap and transmission**



Task Group of AAPM

- TG 50 MLC
- TG 58 EPID
- TG 76 Respiratory Motion
- TG 106 Beam data commissioning equipment and procedures
- TG 104 KV X-ray imaging for patient setup
- TG 100 evaluating QA needs in RT
- TG 148 Tomotherapy QA
- TG 40 COMPREHENSIVE QA FOR RADIATION ONCOLOGY
- TG 135 QA for Robotic Radio surgery
- TG 119** IMRT commissioning: Multiple institution planning and dosimetry comparisons
- TG 147 QA for non-Radiographic RT localization and Positioning
- TG 142** Quality assurance of medical accelerators
- TG 198 Describe the techniques for performing QA test**

QA of IMRT

- Acceptance testing
 - Check whether specifications meet requirements for accurate IMRT and VMAT delivery
- Commissioning
 - Acquisition of the input parameters for the treatment planning system
 - Testing of the dose calculation algorithm
 - E2E testing of the complete treatment chain (audit ?)
- Routine MLC QA
 - E.g. stability of MLC and machine output characteristics
- Patient specific QA
 - Pretreatment measurements and in-vivo dosimetry
 - 3D dose calculation

TG 119

TG 198



Received: 9 February 2021 | Revised: 10 March 2021 | Accepted: 20 April 2021
 DOI: 10.1002/mp.14552

AAPM SCIENTIFIC REPORT **MEDICAL PHYSICS**

AAPM Task Group 198 Report: An implementation guide for TG 142 quality assurance of medical accelerators



Joseph Hanley¹ | Sean Dresser² | William Simon³ | Ryan Flynn⁴ |
 Eric E. Klein⁵ | Daniel Letourneau⁶ | Chihray Liu⁷ | Fang-Fang Yin⁸ |
 Bijan Arjomandy⁹ | Lijun Ma¹⁰ | Francisco Aguirre¹¹ | Jimmy Jones¹² |
 John Bayouth¹³ | Todd Holmes¹⁴



¹Princeton Radiation Oncology, Monroe, New Jersey 08831, USA
²Winship Cancer Institute, Radiation Oncology, Emory University, Atlanta, Georgia 30322, USA
³Royal Nuclear Corp, Melbourne, Florida 32940, USA
⁴Department of Radiation Oncology, University of Iowa, Iowa City, Iowa, 52242, USA


Abstract
 The charges on this task group (TG) were as follows: (a) provide specific procedural guidelines for performing the tests recommended in TG 142; (b) provide estimate of the range of time, appropriate personnel, and qualifications necessary to complete the tests in TG 142, and (c) provide sample daily, weekly, monthly, or annual quality assurance (QA) forms. Many of the guidelines in this report are drawn from the literature and are included in the references. When literature was

IMRT

- Many small segments are made by the MLC for building a dose distribution
- Leaf positioning an travelling becomes important
- Dose calculation are more complex
 - tongue and groves effects
 - Leaf rounding effects
 - Leaf transmission
- Small field Dosimetry becomes important





Dynamic MLC mode technique(s)

Beam stays on during

- Dynamic MLC leaf movement
- Gantry movement
- Dose rate modulation possible
- (couch movement)
- (collimator movement)

- Leaf speeds from 2-7 cm/sec
- Few 100ths segments per beam/arc
- More versatile - more intensity levels
- Could be faster overall treatment time - beam stays on
- The method of choice when combined with gantry rotation (VMAT, ...)


Geometrical accuracy much more important than for conventional treatment


- Position and speed of MLC (IMRT)
- Position and speed of gantry (VMAT)
- Position and speed of table (Tomotherapy)



Important during commissioning AND periodic quality control

IMRT dynamic delivery

- Picket fence test
- Relative position of individual leaves
 - Intentional error of 1 mm in three leaf positions
 - Film measurement









Leaf Positioning

- For 3DCRT the accuracy of the leaf positioning affects the borders of the radiation field. Typically, errors of 1 mm are accepted and can be accounted for in a CTV-to-PTV margin
- For IMRT leaf positioning errors can also impact the dose inside the target
 - E.g. a 1 mm gap error can introduce a dose error of 5%.
 - Overlapping or underlapping of abutting fields lead to hot and cold spots up to 17%/mm

MLC leaf positioning (also during gantry rotation)


"Loss of lateral electron equilibrium in small field segments"

"A drastic change in absorbed dose for small changes in field size occurs when the field size is as small as those encountered in IMRT."

=> tighter tolerances on leaf position/motion


=> also while gantry is rotating and gravity impacts MLC-leaves, during f.e. a VMAT delivery

Picket-fence test



Leaf position accuracy

- <0.5 mm for modern MLC designs
- Measured using Picket-fence tests on film or EPID







Figure 3. Comparison of MLC accuracy positioning between fixed gantry VMAT and IMRT. (a) Picket-fence test on film for leaf pairs 10-12 and 13-15. (b) Signal profile.



VMAT
IMRT

IMRT commissioning: Multiple institution planning and dosimetry comparisons, a report from AAPM Task Group 119

Gary A. Ezzell
Department of Radiation Oncology, Mayo Clinic Scottsdale, 5777 East Mayo Boulevard, MCSB Concourse, Phoenix, Arizona 85054


Jay W. Burmeister
Wayne State University School of Medicine, Karmanos Cancer Center, 4100 John R Street, Detroit, Michigan 48201

Nesrin Dogan
Department of Radiation Oncology, Virginia Commonwealth University, 401 College Street B-129, Richmond, Virginia 23298

Thomas J. LoSasso and James G. Mechalakos
Department of Medical Physics, Memorial Sloan-Kettering Cancer Center, 1275 York Avenue, New York, New York 10065

Dimitris Mihailidis
Department of Radiation Oncology and Medical Physics, Charleston Radiation Therapy, 3100 MacCorkle Avenue Southeast, Charleston, West Virginia 25304

Andrea Molinuevo
RPC, UT MD Anderson Cancer Center, 1515 Holcombe Boulevard, Houston, Texas 77030



گزارش تست های TG-119 مرکز پردیس نور نیلو:

پیشگام تست های فزاد شد:

نتایج	اقدامات
<p>تخت درمان تعریف شد و تست های لازم جهت Verify کردن جنس ماده اضماعی به تخت، طراحی و با استفاده از فانتوم Delta انجام شد. جزئیات و نتایج تست در پیوست شماره 1 آمده است.</p> <p>نامی A نسبت An202 با نتایج مطلوب Pass شدند. پیوست شماره 2.</p> <p>نتایج در پیوست شماره 2 آمده است.</p> <p>نتایج در پیوست شماره 3 آمده است.</p>	<p>تعریف تخت درمان (L20102) با استفاده از CT اسکن تخت در سیستم طراحی درمان</p> <p>صحت نتایج همگام سیستم طراحی درمان توسط تست های سازمان انرژی اتمی (IAEA)</p> <p>انجام تست Proket fence در زاویه گانتری صفر</p> <p style="text-align: right;">تست های TG-119</p>

تست تخت بدون ترتیب طراحی گردید که پس از چهار زاویه مایل خلفی ۰°، ۱۶۰°، ۲۰۰° و ۲۲۰°، فیلدهایی با ابعاد 10x10 cm² با MU=100 پس از آن با استفاده از فانتوم Delta روی دستگاه اجرا گردید و تست با گاما ۹۹.۸ پاسی شد.

پیوست ۲: نتایج تست An202

Case #	Description	Max. point	Field #	Calculation results (Gy)	Measurement results (Gy)	Deviation (%)	Agreement criteria (%)	Pass/Fail
1	Standard SSD, flatland field	3		2.880	2.807	-0.4	2	P
		9		8.740	8.742	-0.1	4	P
		10		1.240	1.208	-1.0	3	F
2	Missing tissue	1		2.880	2.858	-1.8	3	F
3	Blocked corners	3		2.880	2.808	-0.3	3	F
4	Four field box	5	F1	2.880	1.896	-0.5	2	F
			F2	2.880	2.813	-0.7	3	F
			F3	2.880	1.863	-1.9	3	F
			F4	2.880	2.802	-0.1	3	F
		SUM	8.880	7.868	-0.6	3	F	
		8	F1	8.130	8.128	-0.1	4	P
			F2	1.380	1.368	-0.4	2	F
			F3	8.130	8.108	-0.3	4	P
			F4	2.720	2.708	-0.5	3	F
		SUM	4.310	4.303	-0.7	3	F	
		10	F1	1.430	1.408	-1.0	3	F
			F2	8.740	8.138	-0.2	4	F
F3	2.880		2.768	-1.0	3	F		
F4	8.740		8.742	0.0	4	P		
SUM	4.510	4.538	-0.4	3	F			
5	Customized blocking	2		2.880	2.818	-0.3	3	F
7			1.720	1.818	-0.1	4	F	
3			2.880	2.808	-0.4	3	F	
6	L-shaped field	7		1.380	1.302	-0.2	3	F
10			8.120	8.113	-0.1	3	F	
7	Plan with asymmetric fields and wedges	F1	2.880	1.967	-0.7	2	F	
		F2	2.880	2.852	-1.6	4	F	
		F3	2.880	2.343	-0.1	4	F	
		SUM	6.880	6.863	-1.0	3	F	
8	Plan with non-coplanar fields	F1	2.828	1.968	-1.9	3	F	
		F2	2.880	1.978	-1.4	3	F	
		F3	2.880	1.968	-2.0	3	F	
		SUM	8.588	6.818	-1.8	3	F	

